

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003605.D
 Acq On : 23 Jul 2018 15:36
 Operator : JC/SP
 Sample : J4048-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YT-MW-7

Quant Time: Jul 24 10:21:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	244056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	357622	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	337165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	181529	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	150149	51.97	ug/l	0.00
Spiked Amount			Recovery	=	103.94%	
35) Dibromofluoromethane	5.51	113	137791	50.59	ug/l	0.00
Spiked Amount			Recovery	=	101.18%	
50) Toluene-d8	8.72	98	507071	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	11.14	95	170367	45.60	ug/l	0.00
Spiked Amount			Recovery	=	91.20%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
21) trans-1,2-Dichloroethene	3.17	96	816	0.29	ug/l	99
31) Cyclohexane	5.57	56	971	0.22	ug/l #	84
40) Benzene	6.15	78	3014	0.26	ug/l	93
44) Trichloroethene	7.22	130	1550	0.45	ug/l	95
64) Tetrachloroethene	9.34	164	1173	0.29	ug/l #	80
73) Isopropylbenzene	11.02	105	5522	0.44	ug/l	100
83) tert-Butylbenzene	11.77	119	2654	0.26	ug/l	98
85) sec-Butylbenzene	11.95	105	5956	0.49	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003605.D
Acq On : 23 Jul 2018 15:36
Operator : JC/SP
Sample : J4048-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YT-MW-7

Quant Time: Jul 24 10:21:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

