

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003607.D
 Acq On : 23 Jul 2018 16:29
 Operator : JC/SP
 Sample : J4048-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YT-MW-9

Quant Time: Jul 24 10:22:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	245633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	362710	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184968	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	152959	52.61	ug/l	0.00
Spiked Amount			Recovery	=	105.22%	
35) Dibromofluoromethane	5.51	113	139051	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
50) Toluene-d8	8.72	98	514305	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	11.14	95	171989	45.39	ug/l	0.00
Spiked Amount			Recovery	=	90.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003607.D
Acq On : 23 Jul 2018 16:29
Operator : JC/SP
Sample : J4048-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YT-MW-9

Quant Time: Jul 24 10:22:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

