

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003612.D
 Acq On : 23 Jul 2018 18:43
 Operator : JC/SP
 Sample : J4048-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YT-MW-6

Quant Time: Jul 24 08:25:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247922	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	359552	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	333733	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	181520	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	148933	50.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.50%	
35) Dibromofluoromethane	5.51	113	136845	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
50) Toluene-d8	8.72	98	508733	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	11.14	95	170615	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
Target Compounds						
16) Acetone	2.45	43	6034	5.44	ug/l	# 82
19) Methyl tert-butyl Ether	3.21	73	4984	0.60	ug/l	92
21) trans-1,2-Dichloroethene	3.17	96	3210	1.14	ug/l	87
27) cis-1,2-Dichloroethene	4.60	96	36165	11.53	ug/l	98
44) Trichloroethene	7.22	130	9295	2.66	ug/l	98
64) Tetrachloroethene	9.34	164	23704	5.83	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003612.D
Acq On : 23 Jul 2018 18:43
Operator : JC/SP
Sample : J4048-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YT-MW-6

Quant Time: Jul 24 08:25:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

