

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003622.D
 Acq On : 24 Jul 2018 00:29
 Operator : JC/SP
 Sample : J4073-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB02-071718

Quant Time: Aug 02 11:43:59 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254160	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	365841	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	347344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	150082	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
35) Dibromofluoromethane	5.51	113	138768	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
50) Toluene-d8	8.72	98	519180	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
62) 4-Bromofluorobenzene	11.14	95	176838	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
Target Compounds						
16) Acetone	2.45	43	12072	10.62	ug/l	# 85
25) 2-Butanone	4.71	43	5113	2.68	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003622.D
Acq On : 24 Jul 2018 00:29
Operator : JC/SP
Sample : J4073-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB02-071718

Quant Time: Aug 02 11:43:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

