

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003628.D
 Acq On : 24 Jul 2018 03:09
 Operator : JC/SP
 Sample : J4073-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE109D3-071618

Quant Time: Jul 24 09:03:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	359472	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340966	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	190753	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	141524	47.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.68%	
35) Dibromofluoromethane	5.51	113	134509	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
50) Toluene-d8	8.72	98	503866	48.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.92%	
62) 4-Bromofluorobenzene	11.14	95	171650	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	2762	0.99	ug/l	92
12) 1,1-Dichloroethene	2.37	96	943	0.37	ug/l #	53
16) Acetone	2.45	43	7022	6.22	ug/l	93
30) Chloroform	5.22	83	1293	0.26	ug/l #	81
38) Carbon Tetrachloride	5.78	117	1426	0.37	ug/l	96
44) Trichloroethene	7.22	130	134061	38.36	ug/l	96
49) 1,4-Dioxane	7.77	88	900	13.48	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003628.D
Acq On : 24 Jul 2018 03:09
Operator : JC/SP
Sample : J4073-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE109D3-071618

Quant Time: Jul 24 09:03:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

