

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072318\
 Data File : VX003631.D
 Acq On : 24 Jul 2018 04:29
 Operator : JC/SP
 Sample : J4073-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-071718

Quant Time: Aug 01 08:11:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248087	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	355895	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	336257	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186631	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	138817	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
35) Dibromofluoromethane	5.51	113	131875	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
50) Toluene-d8	8.72	98	493907	48.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.96%	
62) 4-Bromofluorobenzene	11.14	95	168780	45.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.38	101	7670	2.787	ug/l	96
12) 1,1-Dichloroethene	2.37	96	10583	4.177	ug/l	80
16) Acetone	2.45	43	5371	4.842	ug/l #	83
24) 1,1-Dichloroethane	3.70	63	20014	4.041	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	18545	5.908	ug/l	72
30) Chloroform	5.22	83	14159	2.928	ug/l	100
32) 1,1,1-Trichloroethane	5.49	97	2680	0.647	ug/l #	48
38) Carbon Tetrachloride	5.78	117	3801	1.008	ug/l	95
44) Trichloroethene	7.22	130	5175978	1496.027	ug/l	98
49) 1,4-Dioxane	7.77	88	926	14.011	ug/l	99
55) 1,1,2-Trichloroethane	9.22	97	4454	1.533	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072318\
Data File : VX003631.D
Acq On : 24 Jul 2018 04:29
Operator : JC/SP
Sample : J4073-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-071718

Quant Time: Aug 01 08:11:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration

