

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003592.D
 Acq On : 23 Jul 2018 09:44
 Operator : JC/SP
 Sample : VX0723WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0723WBL01

Quant Time: Jul 24 06:45:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	299919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387140	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	368978	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200714	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	158417	44.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.24%	
35) Dibromofluoromethane	5.50	113	147205	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
50) Toluene-d8	8.72	98	555125	50.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.18%	
62) 4-Bromofluorobenzene	11.14	95	185543	45.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.74%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003592.D
Acq On : 23 Jul 2018 09:44
Operator : JC/SP
Sample : VX0723WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0723WBL01

Quant Time: Jul 24 06:45:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.01 min

Lab File: VX003592.D

Acq: 23 Jul 2018 09:44

Instrument :

MSVOA_X

ClientSampleId :


VX0723WBL01

Tot Ion:168 Resp: 299919

Ion Ratio Lower Upper

168 100

99 41.5 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

120000

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 44.622 ug/l

RT: 6.06 min Scan# 898

Delta R.T. -0.01 min

Lab File: VX003592.D


Acq: 23 Jul 2018 09:44

Tgt Ion: 65 Resp: 158417

Ion Ratio Lower Upper

65 100

67 53.9 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

80000


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20 6.30





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 1029

Delta R.T. -0.01 min

Lab File: VX003592.D

Acq: 23 Jul 2018 09:44

Instrument :

MSVOA_X

ClientSampleId :

VX0723WBL01


Tot Ion:114 Resp: 387140

Ion Ratio Lower Upper

114 100

63 19.3 0.0 41.4




88 15.1 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.921 ug/l

RT: 5.50 min Scan# 806

Delta R.T. -0.01 min

Lab File: VX003592.D

Acq: 23 Jul 2018 09:44


Tgt Ion:113 Resp: 147205

Ion Ratio Lower Upper

113 100

111 102.2 81.4 122.0



192 25.8 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.087 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003592.D
Acq: 23 Jul 2018 09:44




Instrument :
MSVOA_X
ClientSampleId :
VX0723WBL01


Tot Ion: 98 Resp: 555125
Ion Ratio Lower Upper
98 100
100 64.3 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 45.873 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX003592.D
Acq: 23 Jul 2018 09:44

Tgt Ion: 95 Resp: 185543
Ion Ratio Lower Upper
95 100
174 98.6 0.0 187.4
176 95.3 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX003592.D
Acq: 23 Jul 2018 09:44

Instrument :
MSVOA_X
ClientSampleId :
VX0723WBL01

Tot Ion:117 Resp: 368978
Ion Ratio Lower Upper
117 100
82 52.3 45.0 67.4
119 32.6 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. 0.00 min
Lab File: VX003592.D
Acq: 23 Jul 2018 09:44

Tgt Ion:152 Resp: 200714
Ion Ratio Lower Upper
152 100
115 53.4 40.1 120.2
150 156.1 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

