

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072318\
 Data File : VX003607.D
 Acq On : 23 Jul 2018 16:29
 Operator : JC/SP
 Sample : J4048-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 YT-MW-9

Quant Time: Jul 24 10:22:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 19 17:09:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	245633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	362710	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184968	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	152959	52.61	ug/l	0.00
Spiked Amount			Recovery	=	105.22%	
35) Dibromofluoromethane	5.51	113	139051	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
50) Toluene-d8	8.72	98	514305	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	11.14	95	171989	45.39	ug/l	0.00
Spiked Amount			Recovery	=	90.78%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072318\
Data File : VX003607.D
Acq On : 23 Jul 2018 16:29
Operator : JC/SP
Sample : J4048-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
YT-MW-9

Quant Time: Jul 24 10:22:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 19 17:09:44 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 834

Delta R.T. 0.01 min

Lab File: VX003607.D

Acq: 23 Jul 2018 16:29

Instrument :

MSVOA_X

ClientSampleId :


YT-MW-9

Tot Ion:168 Resp: 245633

Ion Ratio Lower Upper




168 100

99 48.1 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 52.61 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.01 min

Lab File: VX003607.D


Acq: 23 Jul 2018 16:29

Tgt Ion: 65 Resp: 152959

Ion Ratio Lower Upper



65 100


67 53.8 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. -0.00 min

Lab File: VX003607.D

Acq: 23 Jul 2018 16:29

Instrument :

MSVOA_X

ClientSampleId :

YT-MW-9


Tot Ion:114 Resp: 362710

Ion Ratio Lower Upper

114 100

63 19.3 0.0 41.4




88 15.1 0.0 30.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.33 ug/l

RT: 5.51 min Scan# 807

Delta R.T. -0.00 min

Lab File: VX003607.D

Acq: 23 Jul 2018 16:29


Tgt Ion:113 Resp: 139051

Ion Ratio Lower Upper

113 100

111 102.1 81.4 122.0



192 24.5 19.1 28.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.53 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003607.D
Acq: 23 Jul 2018 16:29




Instrument :
MSVOA_X
ClientSampleId :
YT-MW-9


Tot Ion: 98 Resp: 514305
Ion Ratio Lower Upper
98 100
100 63.9 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 45.39 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX003607.D
Acq: 23 Jul 2018 16:29

Tgt Ion: 95 Resp: 171989
Ion Ratio Lower Upper
95 100
174 96.7 0.0 187.4
176 92.5 0.0 181.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX003607.D
Acq: 23 Jul 2018 16:29

Instrument :
MSVOA_X
ClientSampleId :
YT-MW-9

Tot Ion:117 Resp: 340340
Ion Ratio Lower Upper
117 100
82 53.2 45.0 67.4
119 32.2 25.8 38.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1886
Delta R.T. 0.01 min
Lab File: VX003607.D
Acq: 23 Jul 2018 16:29

Tgt Ion:152 Resp: 184968
Ion Ratio Lower Upper
152 100
115 54.5 40.1 120.2
150 156.7 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

