


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX072318\
 Data File : VX003590.D
 Acq On : 23 Jul 2018 08:30
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071918W.M
 Title : SW846 8260
 Last Update : Thu Jul 19 17:09:44 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	3699	PASS
75	95	30	60	44.5	10743	PASS
95	95	100	100	100.0	24133	PASS
96	95	5	9	6.3	1527	PASS
173	174	0.00	2	1.2	280	PASS
174	95	50	100	99.0	23880	PASS
175	174	5	9	7.7	1834	PASS
176	174	95	101	96.9	23128	PASS
177	176	5	9	6.1	1420	PASS