

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX072319\
 Data File : VX011047.D
 Acq On : 23 Jul 2019 09:40
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 24 05:06:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 20 04:05:59 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.467	0.501	-7.3	94	0.00
3 P	Chloromethane	0.460	0.425	7.6	92	0.00
4 C	Vinyl Chloride	0.469	0.443	5.5#	91	0.00
5 T	Bromomethane	0.264	0.229	13.3	92	0.00
6 T	Chloroethane	0.298	0.292	2.0	98	0.00
7 T	Trichlorofluoromethane	0.794	0.777	2.1	97	0.00
8 T	Diethyl Ether	0.274	0.252	8.0	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.453	0.452	0.2	102	0.00
10 T	Methyl Iodide	0.595	0.553	7.1	88	0.00
11 T	Tert butyl alcohol	0.124	0.112	9.7	97	0.00
12 CM	1,1-Dichloroethene	0.445	0.425	4.5#	95	0.00
13 T	Acrolein	0.072	0.060	16.7	87	0.00
14 T	Allyl chloride	0.647	0.618	4.5	96	0.00
15 T	Acrylonitrile	0.250	0.236	5.6	97	0.00
16 T	Acetone	0.240	0.291	-21.2#	123	0.00
17 T	Carbon Disulfide	1.142	1.014	11.2	88	0.00
18 T	Methyl Acetate	0.547	0.524	4.2	104	0.00
19 T	Methyl tert-butyl Ether	1.434	1.392	2.9	100	0.00
20 T	Methylene Chloride	0.509	0.475	6.7	97	0.00
21 T	trans-1,2-Dichloroethene	0.473	0.444	6.1	95	0.00
22 T	Diisopropyl ether	1.364	1.306	4.3	99	0.00
23 T	Vinyl Acetate	1.156	1.140	1.4	96	0.00
24 P	1,1-Dichloroethane	0.792	0.750	5.3	98	0.00
25 T	2-Butanone	0.345	0.363	-5.2	105	0.00
26 T	2,2-Dichloropropane	0.651	0.652	-0.2	102	0.00
27 T	cis-1,2-Dichloroethene	0.547	0.523	4.4	99	0.00
28 T	Bromochloromethane	0.373	0.345	7.5	97	0.00
29 T	Tetrahydrofuran	0.218	0.203	6.9	94	0.00
30 C	Chloroform	0.856	0.814	4.9#	100	0.00
31 T	Cyclohexane	0.722	0.683	5.4	97	0.00
32 T	1,1,1-Trichloroethane	0.741	0.730	1.5	99	0.00
33 S	1,2-Dichloroethane-d4	0.532	0.513	3.6	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.317	0.320	-0.9	99	0.00
36 T	1,1-Dichloropropene	0.411	0.406	1.2	97	0.00
37 T	Ethyl Acetate	0.421	0.413	1.9	93	0.00
38 T	Carbon Tetrachloride	0.428	0.448	-4.7	101	0.00
39 T	Methylcyclohexane	0.533	0.533	0.0	98	0.00
40 TM	Benzene	1.258	1.237	1.7	97	0.00
41 T	Methacrylonitrile	0.239	0.231	3.3	100	0.00
42 TM	1,2-Dichloroethane	0.437	0.436	0.2	101	0.00
43 T	Isopropyl Acetate	0.692	0.692	0.0	98	0.00
44 TM	Trichloroethene	0.372	0.359	3.5	96	0.00
45 C	1,2-Dichloropropane	0.320	0.314	1.9#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX072319\
 Data File : VX011047.D
 Acq On : 23 Jul 2019 09:40
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 24 05:06:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jul 20 04:05:59 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.229	0.229	0.0	98	0.00
47 T	Bromodichloromethane	0.407	0.430	-5.7	101	0.00
48 T	Methyl methacrylate	0.337	0.328	2.7	96	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	95	0.00
50 S	Toluene-d8	1.190	1.192	-0.2	97	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.454	-1.3	97	0.00
52 CM	Toluene	0.826	0.812	1.7#	96	0.00
53 T	t-1,3-Dichloropropene	0.438	0.470	-7.3	100	0.00
54 T	cis-1,3-Dichloropropene	0.487	0.509	-4.5	97	0.00
55 T	1,1,2-Trichloroethane	0.340	0.340	0.0	99	0.00
56 T	Ethyl methacrylate	0.498	0.521	-4.6	98	0.00
57 T	1,3-Dichloropropane	0.540	0.534	1.1	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.238	0.249	-4.6	99	0.00
59 T	2-Hexanone	0.348	0.371	-6.6	100	0.00
60 T	Dibromochloromethane	0.345	0.373	-8.1	99	0.00
61 T	1,2-Dibromoethane	0.361	0.366	-1.4	97	0.00
62 S	4-Bromofluorobenzene	0.438	0.461	-5.3	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.393	0.387	1.5	101	0.00
65 PM	Chlorobenzene	1.027	1.001	2.5	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.370	-2.5	98	0.00
67 C	Ethyl Benzene	1.748	1.752	-0.2#	99	0.00
68 T	m/p-Xylenes	0.671	0.676	-0.7	99	0.00
69 T	o-Xylene	0.654	0.652	0.3	98	0.00
70 T	Styrene	1.091	1.136	-4.1	99	0.00
71 P	Bromoform	0.288	0.314	-9.0	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.378	3.283	2.8	101	0.00
74 T	N-amyl acetate	1.291	1.271	1.5	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.106	1.048	5.2	97	0.00
76 T	1,2,3-Trichloropropane	0.902	0.901	0.1	101	0.00
77 T	Bromobenzene	0.932	0.891	4.4	100	0.00
78 T	n-propylbenzene	3.733	3.748	-0.4	102	0.00
79 T	2-Chlorotoluene	2.242	2.195	2.1	101	0.00
80 T	1,3,5-Trimethylbenzene	2.823	2.793	1.1	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.328	0.326	0.6	98	0.00
82 T	4-Chlorotoluene	2.610	2.608	0.1	103	0.00
83 T	tert-Butylbenzene	2.773	2.760	0.5	101	0.00
84 T	1,2,4-Trimethylbenzene	2.845	2.850	-0.2	101	0.00
85 T	sec-Butylbenzene	3.249	3.333	-2.6	103	0.00
86 T	p-Isopropyltoluene	3.004	3.075	-2.4	102	0.00
87 T	1,3-Dichlorobenzene	1.617	1.618	-0.1	103	0.00
88 T	1,4-Dichlorobenzene	1.659	1.593	4.0	100	0.00
89 T	n-Butylbenzene	2.570	2.708	-5.4	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX072319\
Data File : VX011047.D
Acq On : 23 Jul 2019 09:40
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 24 05:06:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Sat Jul 20 04:05:59 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.474	0.499	-5.3	104	0.00
91 T	1,2-Dichlorobenzene	1.603	1.588	0.9	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.231	0.230	0.4	99	0.00
93 T	1,2,4-Trichlorobenzene	1.086	1.114	-2.6	102	0.00
94 T	Hexachlorobutadiene	0.538	0.549	-2.0	104	0.00
95 T	Naphthalene	3.208	3.291	-2.6	100	0.00
96 T	1,2,3-Trichlorobenzene	1.081	1.107	-2.4	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6