

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017495.D
 Acq On : 23 Jul 2020 16:22
 Operator : JC/SP
 Sample : L3416-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW180D-GW-20200720

Quant Time: Jul 24 06:28:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	503608	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	908925	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	946340	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	498914	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	362499	56.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.08%	
35) Dibromofluoromethane	5.47	113	276179	45.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.86%	
50) Toluene-d8	8.70	98	1124269	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
62) 4-Bromofluorobenzene	11.12	95	485782	55.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.70%	
Target Compounds						
16) Acetone	2.42	43	11861	4.166	ug/l #	83
44) Trichloroethene	7.19	130	28613	3.918	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017495.D
Acq On : 23 Jul 2020 16:22
Operator : JC/SP
Sample : L3416-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW180D-GW-20200720

Quant Time: Jul 24 06:28:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

