

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017501.D
 Acq On : 23 Jul 2020 18:38
 Operator : JC/SP
 Sample : L3186-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-03-072120

Quant Time: Jul 24 06:42:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	482512	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	831667	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	858939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	443784	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	332946	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
35) Dibromofluoromethane	5.47	113	271235	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
50) Toluene-d8	8.70	98	991066	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	11.12	95	433826	54.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.04%	
Target Compounds						
16) Acetone	2.42	43	79028	28.969	ug/l	94
18) Methyl Acetate	2.76	43	7605	1.184	ug/l #	84
20) Methylene Chloride	2.84	84	43553	7.382	ug/l	95
43) Isopropyl Acetate	6.42	43	154151	10.705	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017501.D
Acq On : 23 Jul 2020 18:38
Operator : JC/SP
Sample : L3186-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-03-072120

Quant Time: Jul 24 06:42:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

