

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017510.D
 Acq On : 23 Jul 2020 23:12
 Operator : JC/SP
 Sample : PB130403TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130403TB

Quant Time: Jul 24 07:03:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	482651	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	758496	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	826747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	453868	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	348873	56.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.54%	
35) Dibromofluoromethane	5.47	113	273122	53.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.68%	
50) Toluene-d8	8.70	98	927012	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	11.12	95	425239	58.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.12%	
Target Compounds						
16) Acetone	2.42	43	72039	26.400	ug/l	93
18) Methyl Acetate	2.76	43	6821	1.061	ug/l #	79
20) Methylene Chloride	2.83	84	38743	6.419	ug/l	94
43) Isopropyl Acetate	6.41	43	168469	12.828	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017510.D
Acq On : 23 Jul 2020 23:12
Operator : JC/SP
Sample : PB130403TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130403TB

Quant Time: Jul 24 07:03:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

