

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017515.D
 Acq On : 24 Jul 2020 01:06
 Operator : JC/SP
 Sample : PB130403ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130403ZHE#05

Quant Time: Jul 24 07:12:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	487440	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	802510	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	891825	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	429592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	317937	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
35) Dibromofluoromethane	5.46	113	263147	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	8.70	98	917899	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
62) 4-Bromofluorobenzene	11.12	95	463415	59.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.60%	
Target Compounds						
16) Acetone	2.42	43	68131	24.722	ug/l	97
18) Methyl Acetate	2.75	43	6755	1.041	ug/l #	76
20) Methylene Chloride	2.84	84	49045	8.380	ug/l	92
43) Isopropyl Acetate	6.42	43	153465	11.045	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017515.D
Acq On : 24 Jul 2020 01:06
Operator : JC/SP
Sample : PB130403ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130403ZHE#05

Quant Time: Jul 24 07:12:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

