

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017517.D
 Acq On : 24 Jul 2020 01:51
 Operator : JC/SP
 Sample : PB130403ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130403ZHE#07

Quant Time: Jul 24 07:16:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	490223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	846509	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	934583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	487409	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	346257	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
35) Dibromofluoromethane	5.46	113	283017	49.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.98%	
50) Toluene-d8	8.70	98	1076707	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
62) 4-Bromofluorobenzene	11.12	95	476452	58.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.58%	
Target Compounds						
16) Acetone	2.42	43	69638	25.126	ug/l	99
18) Methyl Acetate	2.75	43	7356	1.127	ug/l #	79
20) Methylene Chloride	2.84	84	47402	8.002	ug/l	94
43) Isopropyl Acetate	6.41	43	160965	10.982	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017517.D
Acq On : 24 Jul 2020 01:51
Operator : JC/SP
Sample : PB130403ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130403ZHE#07

Quant Time: Jul 24 07:16:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

