

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017519.D
 Acq On : 24 Jul 2020 02:37
 Operator : JC/SP
 Sample : PB130403ZHE#09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130403ZHE#09

Quant Time: Jul 24 07:19:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	465807	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	849913	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	810324	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	410624	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	321548	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
35) Dibromofluoromethane	5.46	113	262132	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
50) Toluene-d8	8.70	98	951871	46.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.84%	
62) 4-Bromofluorobenzene	11.12	95	409570	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
Target Compounds						
16) Acetone	2.42	43	71187	27.031	ug/l	98
18) Methyl Acetate	2.75	43	7724	1.245	ug/l	92
20) Methylene Chloride	2.84	84	41629	7.296	ug/l	94
43) Isopropyl Acetate	6.42	43	171460	11.652	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017519.D
Acq On : 24 Jul 2020 02:37
Operator : JC/SP
Sample : PB130403ZHE#09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130403ZHE#09

Quant Time: Jul 24 07:19:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

