

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\  
 Data File : VX017523.D  
 Acq On : 24 Jul 2020 04:08  
 Operator : JC/SP  
 Sample : PB130403ZHE#13  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 48 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PB130403ZHE#13

Quant Time: Jul 24 07:32:46 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X072320W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 24 05:48:09 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 5.63  | 168  | 474319   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.84  | 114  | 817728   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 915460   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 485612   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.04  | 65   | 330134   | 54.18  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 108.36% |          |
| 35) Dibromofluoromethane    | 5.47  | 113  | 277686   | 50.77  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 101.54% |          |
| 50) Toluene-d8              | 8.70  | 98   | 1050385  | 53.24  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 106.48% |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 465041   | 58.90  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 117.80% |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 2.42  | 43   | 75387    | 28.112 | ug/l    | 94       |
| 18) Methyl Acetate          | 2.75  | 43   | 7825     | 1.239  | ug/l #  | 84       |
| 20) Methylene Chloride      | 2.84  | 84   | 44727    | 7.771  | ug/l    | 97       |
| 43) Isopropyl Acetate       | 6.42  | 43   | 160330   | 11.324 | ug/l    | 100      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\  
Data File : VX017523.D  
Acq On : 24 Jul 2020 04:08  
Operator : JC/SP  
Sample : PB130403ZHE#13  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 48 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PB130403ZHE#13

Quant Time: Jul 24 07:32:46 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X072320W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 24 05:48:09 2020  
Response via : Initial Calibration

