

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017526.D
 Acq On : 24 Jul 2020 05:16
 Operator : JC/SP
 Sample : L3380-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-241

Quant Time: Jul 24 07:38:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	480500	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	825278	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	921281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	480979	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	341291	55.29	ug/l	0.00
Spiked Amount						
			Recovery	=	110.58%	
35) Dibromofluoromethane	5.46	113	283970	51.45	ug/l	0.00
Spiked Amount						
			Recovery	=	102.90%	
50) Toluene-d8	8.70	98	1059405	53.21	ug/l	0.00
Spiked Amount						
			Recovery	=	106.42%	
62) 4-Bromofluorobenzene	11.12	95	474783	59.58	ug/l	0.00
Spiked Amount						
			Recovery	=	119.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	10380	7.979	ug/l #	77
16) Acetone	2.42	43	93927	34.575	ug/l	99
18) Methyl Acetate	2.75	43	8652	1.352	ug/l #	86
20) Methylene Chloride	2.84	84	104249	19.595	ug/l	99
25) 2-Butanone	4.64	43	43414	11.249	ug/l	94
43) Isopropyl Acetate	6.42	43	139910	9.791	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017526.D
Acq On : 24 Jul 2020 05:16
Operator : JC/SP
Sample : L3380-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-241

Quant Time: Jul 24 07:38:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

