

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072320\
 Data File : VX017527.D
 Acq On : 24 Jul 2020 05:39
 Operator : JC/SP
 Sample : L3380-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VNJ-242

Quant Time: Jul 24 07:44:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	459099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	791205	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	855826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	459573	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	327454	55.53	ug/l	0.00
Spiked Amount						
			Recovery	=	111.06%	
35) Dibromofluoromethane	5.47	113	271975	51.40	ug/l	0.00
Spiked Amount						
			Recovery	=	102.80%	
50) Toluene-d8	8.70	98	1019684	53.42	ug/l	0.00
Spiked Amount						
			Recovery	=	106.84%	
62) 4-Bromofluorobenzene	11.12	95	444558	58.19	ug/l	0.00
Spiked Amount						
			Recovery	=	116.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	11371	9.149	ug/l	97
16) Acetone	2.42	43	100052	38.546	ug/l	99
18) Methyl Acetate	2.75	43	7053	1.154	ug/l #	75
20) Methylene Chloride	2.84	84	112634	22.330	ug/l	96
25) 2-Butanone	4.64	43	44646	12.108	ug/l	99
43) Isopropyl Acetate	6.42	43	147483	10.766	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072320\
Data File : VX017527.D
Acq On : 24 Jul 2020 05:39
Operator : JC/SP
Sample : L3380-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VNJ-242

Quant Time: Jul 24 07:44:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

