

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023413.D
 Acq On : 23 Jul 2021 10:14
 Operator : JC/MD
 Sample : VX0723MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0723MBL01

Quant Time: Jul 24 03:43:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	185608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	313300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99425	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	118079	51.000	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.000%
35) Dibromofluoromethane	5.397	113	98331	50.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.400%
50) Toluene-d8	8.653	98	363883	50.560	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.085	95	110513	45.158	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023413.D
Acq On : 23 Jul 2021 10:14
Operator : JC/MD
Sample : VX0723MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01

Quant Time: Jul 24 03:43:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

