

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023414.D
 Acq On : 23 Jul 2021 10:37
 Operator : JC/MD
 Sample : VX0723WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0723WBL01

Quant Time: Jul 24 03:43:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	306827	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	264594	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107278	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	117593	51.480	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.960%
35) Dibromofluoromethane	5.397	113	97097	50.613	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	8.653	98	360926	51.207	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.420%
62) 4-Bromofluorobenzene	11.085	95	115234	48.080	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023414.D
Acq On : 23 Jul 2021 10:37
Operator : JC/MD
Sample : VX0723WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0723WBL01

Quant Time: Jul 24 03:43:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

