

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023416.D
 Acq On : 23 Jul 2021 11:29
 Operator : JC/MD
 Sample : M3145-01 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-5-PS

Quant Time: Jul 24 03:43:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	183087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	303802	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	264648	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115840	50.721	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.440%
35) Dibromofluoromethane	5.397	113	97339	51.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.480%
50) Toluene-d8	8.653	98	365039	52.307	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.620%
62) 4-Bromofluorobenzene	11.085	95	110349	46.500	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.000%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.800	84	1150	0.528	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	1182	0.533	ug/l	93
64) Tetrachloroethene	9.275	164	579	0.275	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023416.D
Acq On : 23 Jul 2021 11:29
Operator : JC/MD
Sample : M3145-01 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-5-PS

Quant Time: Jul 24 03:43:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

