

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023419.D
 Acq On : 23 Jul 2021 12:40
 Operator : JC/MD
 Sample : M3145-05 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-7-PS

Quant Time: Jul 24 03:44:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	171721	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	283968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	240620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109755	51.238	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.480%	
35) Dibromofluoromethane	5.397	113	90909	51.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.400%	
50) Toluene-d8	8.653	98	337084	51.675	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.340%	
62) 4-Bromofluorobenzene	11.085	95	99179	44.712	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.420%	
Target Compounds						Qvalue
20) Methylene Chloride	2.806	84	1247	0.610	ug/l #	54
27) cis-1,2-Dichloroethene	4.507	96	1146	0.551	ug/l #	33

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023419.D
Acq On : 23 Jul 2021 12:40
Operator : JC/MD
Sample : M3145-05 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-7-PS

Quant Time: Jul 24 03:44:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

