

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023420.D
 Acq On : 23 Jul 2021 13:03
 Operator : JC/MD
 Sample : M3145-09 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-8-PD

Quant Time: Jul 24 03:44:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	171254	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	290608	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92943	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112664	52.739	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.480%	
35) Dibromofluoromethane	5.397	113	92446	50.878	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.760%	
50) Toluene-d8	8.653	98	342986	51.378	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.760%	
62) 4-Bromofluorobenzene	11.085	95	105444	46.451	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.900%	
Target Compounds						Qvalue
20) Methylene Chloride	2.788	84	1348	0.661	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	1323	0.637	ug/l #	19

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023420.D
Acq On : 23 Jul 2021 13:03
Operator : JC/MD
Sample : M3145-09 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-8-PD

Quant Time: Jul 24 03:44:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

