

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023426.D
 Acq On : 23 Jul 2021 16:12
 Operator : JC/MD
 Sample : M3146-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-6

Quant Time: Jul 24 03:45:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	282252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93439	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107631	51.583	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 103.160%		
35) Dibromofluoromethane	5.397	113	89091	50.483	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.960%		
50) Toluene-d8	8.653	98	340894	52.576	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 105.160%		
62) 4-Bromofluorobenzene	11.085	95	106083	48.116	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.240%		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2283	2.230	ug/l #	81
20) Methylene Chloride	2.800	84	1158	0.582	ug/l #	81
44) Trichloroethene	7.135	130	2304	1.108	ug/l	86
52) Toluene	8.726	92	1300	0.277	ug/l	94
64) Tetrachloroethene	9.275	164	4335	2.196	ug/l	95
85) sec-Butylbenzene	11.890	105	9273	1.270	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023426.D
Acq On : 23 Jul 2021 16:12
Operator : JC/MD
Sample : M3146-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-6

Quant Time: Jul 24 03:45:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

