

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023427.D
 Acq On : 23 Jul 2021 16:35
 Operator : JC/MD
 Sample : M3146-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-1S(ONSITE)

Quant Time: Jul 24 03:45:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167119	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	283243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	239569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	107885	51.752	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.500%
35) Dibromofluoromethane	5.397	113	90027	50.835	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.680%
50) Toluene-d8	8.653	98	331426	50.937	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.880%
62) 4-Bromofluorobenzene	11.085	95	102681	46.410	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.820%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	2784	2.722	ug/l	98
64) Tetrachloroethene	9.275	164	817	0.428	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023427.D
Acq On : 23 Jul 2021 16:35
Operator : JC/MD
Sample : M3146-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-1S(ONSITE)

Quant Time: Jul 24 03:45:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

