

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023428.D
 Acq On : 23 Jul 2021 16:59
 Operator : JC/MD
 Sample : M3146-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-6S

Quant Time: Jul 24 03:45:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	178932	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	298265	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	271075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	114663	51.372	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.740%	
35) Dibromofluoromethane	5.403	113	94917	50.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.800%	
50) Toluene-d8	8.653	98	361587	52.774	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.540%	
62) 4-Bromofluorobenzene	11.085	95	116377	49.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.900%	
Target Compounds						Qvalue
16) Acetone	2.392	43	3459	3.159	ug/l	92
20) Methylene Chloride	2.794	84	857	0.402	ug/l	91
64) Tetrachloroethene	9.275	164	1410	0.654	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023428.D
Acq On : 23 Jul 2021 16:59
Operator : JC/MD
Sample : M3146-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-6S

Quant Time: Jul 24 03:45:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

