

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023429.D
 Acq On : 23 Jul 2021 17:22
 Operator : JC/MD
 Sample : M3146-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-MW-10(ONSITE)

Quant Time: Jul 24 03:45:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	158760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	270194	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	234410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82964	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	104288	52.660	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 105.320%		
35) Dibromofluoromethane	5.397	113	84993	50.310	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.620%		
50) Toluene-d8	8.653	98	322225	51.915	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.820%		
62) 4-Bromofluorobenzene	11.085	95	97736	46.308	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.620%		
Target Compounds						Qvalue
16) Acetone	2.386	43	3569	3.674	ug/l	92
20) Methylene Chloride	2.800	84	901	0.477	ug/l #	74
27) cis-1,2-Dichloroethene	4.507	96	970	0.504	ug/l	98
44) Trichloroethene	7.129	130	593	0.298	ug/l	99
64) Tetrachloroethene	9.281	164	1691	0.906	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023429.D
Acq On : 23 Jul 2021 17:22
Operator : JC/MD
Sample : M3146-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-MW-10(ONSITE)

Quant Time: Jul 24 03:45:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

