

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023430.D
 Acq On : 23 Jul 2021 17:46
 Operator : JC/MD
 Sample : M3145-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 152140-DDC-5-PS

Quant Time: Jul 24 03:46:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	168185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	286317	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110993	52.905	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.820%	
35) Dibromofluoromethane	5.397	113	91127	50.904	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.800%	
50) Toluene-d8	8.653	98	335044	50.940	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.880%	
62) 4-Bromofluorobenzene	11.085	95	104145	46.566	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.140%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2531	2.459	ug/l	95
20) Methylene Chloride	2.794	84	1043	0.521	ug/l	98
27) cis-1,2-Dichloroethene	4.507	96	9947	4.879	ug/l	90
30) Chloroform	5.111	83	1203	0.361	ug/l	95
64) Tetrachloroethene	9.281	164	3695	1.879	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023430.D
Acq On : 23 Jul 2021 17:46
Operator : JC/MD
Sample : M3145-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
152140-DDC-5-PS

Quant Time: Jul 24 03:46:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

