

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
 Data File : VX023413.D
 Acq On : 23 Jul 2021 10:14
 Operator : JC/MD
 Sample : VX0723MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0723MBL01

Quant Time: Jul 24 03:43:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	185608	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	313300	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256584	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99425	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	118079	51.000	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.000%
35) Dibromofluoromethane	5.397	113	98331	50.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.400%
50) Toluene-d8	8.653	98	363883	50.560	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.120%
62) 4-Bromofluorobenzene	11.085	95	110513	45.158	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.320%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072321\
Data File : VX023413.D
Acq On : 23 Jul 2021 10:14
Operator : JC/MD
Sample : VX0723MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01

Quant Time: Jul 24 03:43:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.006 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14




Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01


Tgt Ion:168 Resp: 185608
Ion Ratio Lower Upper
168 100
99 57.2 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 51.000 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14

Tgt Ion: 65 Resp: 118079
Ion Ratio Lower Upper
65 100
67 51.4 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. -0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14




Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01


Tgt Ion:114 Resp: 313300
Ion Ratio Lower Upper
114 100
63 19.1 0.0 40.8
88 15.3 0.0 32.8



#35
Dibromofluoromethane
Concen: 50.197 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14

Tgt Ion:113 Resp: 98331
Ion Ratio Lower Upper
113 100
111 102.3 81.9 122.9
192 19.4 16.7 25.1









#50
Toluene-d8
Concen: 50.560 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14




Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01


Tgt Ion: 98 Resp: 363883
Ion Ratio Lower Upper
98 100
100 64.1 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 45.158 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14

Tgt Ion: 95 Resp: 110513
Ion Ratio Lower Upper
95 100
174 80.7 0.0 154.6
176 77.9 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14

Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01

Tgt Ion:117 Resp: 256584
Ion Ratio Lower Upper
117 100
82 56.6 47.4 71.2
119 32.0 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX023413.D
Acq: 23 Jul 2021 10:14

Tgt Ion:152 Resp: 99425
Ion Ratio Lower Upper
152 100
115 57.8 41.1 123.3
150 154.6 0.0 349.0

