

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072325\
 Data File : VX047078.D
 Acq On : 23 Jul 2025 09:58
 Operator : JC/MD
 Sample : VX0723MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0723MBL01

Quant Time: Jul 24 05:43:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	263409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	478754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	452406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	220696	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	191586	58.388	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.780%
35) Dibromofluoromethane	5.391	113	158439	53.599	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.200%
50) Toluene-d8	8.647	98	579378	60.523	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	121.040%#
62) 4-Bromofluorobenzene	11.079	95	232378	56.329	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072325\
Data File : VX047078.D
Acq On : 23 Jul 2025 09:58
Operator : JC/MD
Sample : VX0723MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0723MBL01

Quant Time: Jul 24 05:43:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

