

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072325\
 Data File : VX047085.D
 Acq On : 23 Jul 2025 13:04
 Operator : JC/MD
 Sample : Q2668-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1

Quant Time: Jul 24 05:48:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	366953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	672516	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	644011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	316311	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	241280	52.784	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.560%			
35) Dibromofluoromethane	5.397	113	196015	47.206	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 94.420%			
50) Toluene-d8	8.653	98	711982	52.946	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 105.900%			
62) 4-Bromofluorobenzene	11.079	95	294116	50.753	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.500%			
Target Compounds						
16) Acetone	2.392	43	53372	29.146	ug/l	99
18) Methyl Acetate	2.715	43	32249	6.627	ug/l	98
20) Methylene Chloride	2.806	84	18804	3.733	ug/l	97
37) Ethyl Acetate	4.751	43	18530m	2.841	ug/l	
40) Benzene	6.056	78	35851	1.750	ug/l	97
43) Isopropyl Acetate	6.348	43	12979	1.276	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072325\
Data File : VX047085.D
Acq On : 23 Jul 2025 13:04
Operator : JC/MD
Sample : Q2668-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1

Quant Time: Jul 24 05:48:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

