

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072325\  
 Data File : VX047097.D  
 Acq On : 23 Jul 2025 17:15  
 Operator : JC/MD  
 Sample : Q2481-12 100X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 CC0627-AL

Quant Time: Jul 24 05:54:56 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072125W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jul 22 03:59:51 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.562  | 168   | 551022   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769  | 114   | 978806   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 910956   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 450205   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.964  | 65    | 318125   | 46.347   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 92.700%  |
| 35) Dibromofluoromethane    | 5.397  | 113   | 264258   | 43.726   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 87.460%  |
| 50) Toluene-d8              | 8.647  | 98    | 960520   | 49.077   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 98.160%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 387800   | 45.979   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 91.960%  |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX072325\  
Data File : VX047097.D  
Acq On : 23 Jul 2025 17:15  
Operator : JC/MD  
Sample : Q2481-12 100X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
CC0627-AL

Quant Time: Jul 24 05:54:56 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072125W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jul 22 03:59:51 2025  
Response via : Initial Calibration

