

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072325\
 Data File : VX047099.D
 Acq On : 23 Jul 2025 17:57
 Operator : JC/MD
 Sample : Q2481-19 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CC0627-CLOXAL

Quant Time: Jul 24 05:55:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	488528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	872023	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	813709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	404152	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	317621	52.193	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.380%	
35) Dibromofluoromethane	5.397	113	260814	48.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.880%	
50) Toluene-d8	8.647	98	941158	53.976	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.960%	
62) 4-Bromofluorobenzene	11.079	95	389770	51.871	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.740%	
Target Compounds						Qvalue
20) Methylene Chloride	2.806	84	8846	1.319	ug/l	96
71) Bromoform	10.799	173	6944	1.424	ug/l #	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072325\
Data File : VX047099.D
Acq On : 23 Jul 2025 17:57
Operator : JC/MD
Sample : Q2481-19 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CC0627-CLOXAL

Quant Time: Jul 24 05:55:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

