


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX072418\  
Data File : VX003641.D  
Acq On : 24 Jul 2018 15:40  
Operator : JC/SP  
Sample : BFB  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X072418W.M  
Title : SW846 8260  
Last Update : Wed Jul 25 04:33:52 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1724

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 50          | 95           | 15           | 40           | 18.6      | 2363    | PASS             |
| 75          | 95           | 30           | 60           | 49.2      | 6250    | PASS             |
| 95          | 95           | 100          | 100          | 100.0     | 12706   | PASS             |
| 96          | 95           | 5            | 9            | 7.1       | 908     | PASS             |
| 173         | 174          | 0.00         | 2            | 1.0       | 118     | PASS             |
| 174         | 95           | 50           | 100          | 88.6      | 11257   | PASS             |
| 175         | 174          | 5            | 9            | 7.9       | 889     | PASS             |
| 176         | 174          | 95           | 101          | 98.0      | 11031   | PASS             |
| 177         | 176          | 5            | 9            | 6.4       | 702     | PASS             |