

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017536.D
 Acq On : 24 Jul 2020 12:42
 Operator : JC/SP
 Sample : L3416-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW180D2-GW-20200722DL

Quant Time: Jul 25 02:27:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	425821	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	683987	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	742239	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	376121	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	276195	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
35) Dibromofluoromethane	5.47	113	234309	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
50) Toluene-d8	8.70	98	878159	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
62) 4-Bromofluorobenzene	11.12	95	369731	55.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.96%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.35	101	19033	4.115	ug/l	95
12) 1,1-Dichloroethene	2.35	96	10049	2.285	ug/l	83
27) cis-1,2-Dichloroethene	4.56	96	3604	0.704	ug/l	39
44) Trichloroethene	7.18	130	227915	41.467	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017536.D
Acq On : 24 Jul 2020 12:42
Operator : JC/SP
Sample : L3416-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW180D2-GW-20200722DL

Quant Time: Jul 25 02:27:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

