

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017537.D
 Acq On : 24 Jul 2020 13:05
 Operator : JC/SP
 Sample : L3416-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW180D1-GW-20200720DL

Quant Time: Jul 25 02:29:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	452656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	724905	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	783357	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	403935	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	303861	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
35) Dibromofluoromethane	5.47	113	247038	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
50) Toluene-d8	8.70	98	899688	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.12	95	375004	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	3260	0.663	ug/l #	79
44) Trichloroethene	7.19	130	269534	46.271	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017537.D
Acq On : 24 Jul 2020 13:05
Operator : JC/SP
Sample : L3416-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW180D1-GW-20200720DL

Quant Time: Jul 25 02:29:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

