

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017551.D
 Acq On : 24 Jul 2020 18:25
 Operator : JC/SP
 Sample : L3381-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 365-GRID-B6-B8

Quant Time: Jul 25 03:45:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	452640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	743534	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	812752	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	456192	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	312430	53.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.46%	
35) Dibromofluoromethane	5.47	113	269124	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
50) Toluene-d8	8.70	98	923985	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
62) 4-Bromofluorobenzene	11.12	95	403047	56.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.28%	
Target Compounds						
16) Acetone	2.42	43	44833	17.519	ug/l	97
18) Methyl Acetate	2.75	43	6903	1.145	ug/l #	78
20) Methylene Chloride	2.83	84	41684	7.559	ug/l	96
43) Isopropyl Acetate	6.42	43	149160	11.586	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017551.D
Acq On : 24 Jul 2020 18:25
Operator : JC/SP
Sample : L3381-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
365-GRID-B6-B8

Quant Time: Jul 25 03:45:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

