

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017553.D
 Acq On : 24 Jul 2020 19:10
 Operator : JC/SP
 Sample : L3379-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-361

Quant Time: Jul 25 03:49:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	453144	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	780972	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	802348	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	439465	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	317107	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	5.47	113	254304	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	8.70	98	949906	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	11.12	95	408249	54.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.28%	
Target Compounds						
16) Acetone	2.42	43	98437	38.423	ug/l	98
18) Methyl Acetate	2.75	43	8442	1.399	ug/l	95
20) Methylene Chloride	2.84	84	43153	7.861	ug/l #	90
25) 2-Butanone	4.64	43	18514	5.087	ug/l	95
43) Isopropyl Acetate	6.42	43	94680	7.002	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017553.D
Acq On : 24 Jul 2020 19:10
Operator : JC/SP
Sample : L3379-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-361

Quant Time: Jul 25 03:49:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

