

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017555.D
 Acq On : 24 Jul 2020 19:56
 Operator : JC/SP
 Sample : L3402-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUS-WASH

Manual Integrations
 APPROVED

MMDadoda
 7/27/2020 11:03:29 AM

Quant Time: Jul 25 03:58:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	496278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	785791	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	890251	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	514247	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	315589	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
35) Dibromofluoromethane	5.46	113	286383	54.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.98%	
50) Toluene-d8	8.70	98	1025834	54.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.22%	
62) 4-Bromofluorobenzene	11.12	95	428027	56.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.82%	
Target Compounds						
16) Acetone	2.42	43	249933	89.077	ug/l	99
20) Methylene Chloride	2.84	84	43862	7.201	ug/l	90
25) 2-Butanone	4.64	43	37873	9.502	ug/l	94
43) Isopropyl Acetate	6.42	43	63447	4.663	ug/l	98
51) 4-Methyl-2-Pentanone	8.62	43	129177m	15.913	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017555.D
Acq On : 24 Jul 2020 19:56
Operator : JC/SP
Sample : L3402-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BUS-WASH

Manual Integrations
APPROVED

MMDadoda
7/27/2020 11:03:29 AM

Quant Time: Jul 25 03:58:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

