

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017566.D
 Acq On : 25 Jul 2020 03:28
 Operator : JC/SP
 Sample : PB130430ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#02

Quant Time: Jul 25 04:41:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	533112	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	885615	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	987182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	515589	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	350366	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
35) Dibromofluoromethane	5.46	113	304917	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
50) Toluene-d8	8.70	98	1137411	53.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.46%	
62) 4-Bromofluorobenzene	11.12	95	500403	58.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.04%	
Target Compounds						
16) Acetone	2.42	43	82487	27.367	ug/l	95
18) Methyl Acetate	2.75	43	8355	1.177	ug/l	91
20) Methylene Chloride	2.84	84	46920	7.165	ug/l	94
43) Isopropyl Acetate	6.41	43	196643	12.824	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017566.D
Acq On : 25 Jul 2020 03:28
Operator : JC/SP
Sample : PB130430ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#02

Quant Time: Jul 25 04:41:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

