

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017568.D
 Acq On : 25 Jul 2020 04:13
 Operator : JC/SP
 Sample : PB130430ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#04

Quant Time: Jul 25 04:58:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	508502	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	891121	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	901227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	492392	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	349443	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
35) Dibromofluoromethane	5.47	113	300084	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	8.70	98	1122577	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
62) 4-Bromofluorobenzene	11.12	95	491582	57.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.26%	
Target Compounds						
16) Acetone	2.42	43	84390	29.354	ug/l	98
18) Methyl Acetate	2.76	43	8022	1.185	ug/l #	78
20) Methylene Chloride	2.84	84	50900	8.330	ug/l	92
43) Isopropyl Acetate	6.42	43	195449	12.668	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017568.D
Acq On : 25 Jul 2020 04:13
Operator : JC/SP
Sample : PB130430ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#04

Quant Time: Jul 25 04:58:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

