

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017569.D
 Acq On : 25 Jul 2020 04:36
 Operator : JC/SP
 Sample : PB130430ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#05

Quant Time: Jul 25 05:02:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	526738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	912223	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	937903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	491950	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	361148	53.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.76%	
35) Dibromofluoromethane	5.47	113	291985	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
50) Toluene-d8	8.70	98	1110218	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	11.12	95	475322	53.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.92%	
Target Compounds						
16) Acetone	2.42	43	45508	15.281	ug/l	92
18) Methyl Acetate	2.75	43	8358	1.192	ug/l #	86
20) Methylene Chloride	2.83	84	20185	2.376	ug/l	93
43) Isopropyl Acetate	6.41	43	176293	11.162	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017569.D
Acq On : 25 Jul 2020 04:36
Operator : JC/SP
Sample : PB130430ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#05

Quant Time: Jul 25 05:02:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

