

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017570.D
 Acq On : 25 Jul 2020 04:59
 Operator : JC/SP
 Sample : PB130430ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#06

Quant Time: Jul 25 06:53:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	513366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	910183	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	834348	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	531822	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	350908	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
35) Dibromofluoromethane	5.46	113	283528	46.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.16%	
50) Toluene-d8	8.70	98	1095573	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
62) 4-Bromofluorobenzene	11.12	95	494782	56.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.60%	
Target Compounds						
16) Acetone	2.42	43	44125	15.203	ug/l	95
18) Methyl Acetate	2.75	43	8852	1.295	ug/l #	78
20) Methylene Chloride	2.84	84	21557	2.730	ug/l	97
43) Isopropyl Acetate	6.42	43	179118	11.366	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017570.D
Acq On : 25 Jul 2020 04:59
Operator : JC/SP
Sample : PB130430ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#06

Quant Time: Jul 25 06:53:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

