

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017572.D
 Acq On : 25 Jul 2020 05:44
 Operator : JC/SP
 Sample : PB130430ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#08

Quant Time: Jul 25 06:56:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	484898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	849858	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	904139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	486237	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	336372	54.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.00%	
35) Dibromofluoromethane	5.46	113	286647	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	8.70	98	1072119	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
62) 4-Bromofluorobenzene	11.12	95	470366	57.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.64%	
Target Compounds						
16) Acetone	2.42	43	43785	15.971	ug/l	94
18) Methyl Acetate	2.75	43	7411	1.148	ug/l #	87
20) Methylene Chloride	2.83	84	21895	3.034	ug/l	97
43) Isopropyl Acetate	6.42	43	183028	12.438	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017572.D
Acq On : 25 Jul 2020 05:44
Operator : JC/SP
Sample : PB130430ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#08

Quant Time: Jul 25 06:56:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

