

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017574.D
 Acq On : 25 Jul 2020 06:30
 Operator : JC/SP
 Sample : PB130430ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#10

Manual Integrations
 APPROVED

MMDadoda
 7/27/2020 11:02:37 AM

Quant Time: Jul 27 07:41:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	488524	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	832732	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	874597	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	430231	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	341751	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
35) Dibromofluoromethane	5.47	113	286732	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
50) Toluene-d8	8.70	98	1036663	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
62) 4-Bromofluorobenzene	11.12	95	411981	51.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.48%	
Target Compounds						
16) Acetone	2.42	43	42311	15.319	ug/l	100
18) Methyl Acetate	2.75	43	6812m	1.047	ug/l	
20) Methylene Chloride	2.84	84	26023	3.817	ug/l	96
43) Isopropyl Acetate	6.42	43	175694	12.186	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017574.D
Acq On : 25 Jul 2020 06:30
Operator : JC/SP
Sample : PB130430ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#10

Manual Integrations
APPROVED

MMDadoda
7/27/2020 11:02:37 AM

Quant Time: Jul 27 07:41:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

