

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017576.D
 Acq On : 25 Jul 2020 07:16
 Operator : JC/SP
 Sample : PB130430ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#12

Quant Time: Jul 27 01:29:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	418416	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	721222	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	821541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	432739	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	329663	61.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.68%	
35) Dibromofluoromethane	5.47	113	269504	55.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.74%	
50) Toluene-d8	8.70	98	1068808	61.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.84%	
62) 4-Bromofluorobenzene	11.12	95	423578	60.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.64%	
Target Compounds						
16) Acetone	2.42	43	44683	18.889	ug/l	99
18) Methyl Acetate	2.75	43	7733	1.388	ug/l #	82
20) Methylene Chloride	2.84	84	19818	3.247	ug/l	94
43) Isopropyl Acetate	6.42	43	166226	13.311	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017576.D
Acq On : 25 Jul 2020 07:16
Operator : JC/SP
Sample : PB130430ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#12

Quant Time: Jul 27 01:29:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

