

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017578.D
 Acq On : 25 Jul 2020 08:02
 Operator : JC/SP
 Sample : PB130430ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#14

Quant Time: Jul 27 01:34:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	416528	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	728167	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	744424	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	411073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	302427	56.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.04%	
35) Dibromofluoromethane	5.47	113	255140	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
50) Toluene-d8	8.70	98	910363	51.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.64%	
62) 4-Bromofluorobenzene	11.12	95	409809	58.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.58%	
Target Compounds						
16) Acetone	2.42	43	76779	32.603	ug/l	96
18) Methyl Acetate	2.75	43	8481	1.529	ug/l #	84
20) Methylene Chloride	2.84	84	51774	10.663	ug/l	93
43) Isopropyl Acetate	6.41	43	169401	13.436	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017578.D
Acq On : 25 Jul 2020 08:02
Operator : JC/SP
Sample : PB130430ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#14

Quant Time: Jul 27 01:34:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

