

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017579.D
 Acq On : 25 Jul 2020 08:24
 Operator : JC/SP
 Sample : PB130430ZHE#15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB130430ZHE#15

Quant Time: Jul 27 01:37:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	466309	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	721540	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	748516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	411264	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	321605	53.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.38%	
35) Dibromofluoromethane	5.47	113	268578	55.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.32%	
50) Toluene-d8	8.70	98	883861	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	11.12	95	397896	57.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.22%	
Target Compounds						
16) Acetone	2.42	43	74969	28.436	ug/l	94
18) Methyl Acetate	2.75	43	6684	1.076	ug/l #	68
20) Methylene Chloride	2.84	84	49706	8.957	ug/l	99
43) Isopropyl Acetate	6.42	43	163088	13.054	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017579.D
Acq On : 25 Jul 2020 08:24
Operator : JC/SP
Sample : PB130430ZHE#15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB130430ZHE#15

Quant Time: Jul 27 01:37:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

