

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017593.D
 Acq On : 25 Jul 2020 14:06
 Operator : JC/SP
 Sample : L3371-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS038-200720

Quant Time: Jul 27 07:31:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	438574	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	742439	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	795503	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	430635	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	305469	54.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.44%	
35) Dibromofluoromethane	5.47	113	256820	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
50) Toluene-d8	8.70	98	936489	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
62) 4-Bromofluorobenzene	11.12	95	407044	56.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.56%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	69247	13.137	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017593.D
Acq On : 25 Jul 2020 14:06
Operator : JC/SP
Sample : L3371-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 67 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS038-200720

Quant Time: Jul 27 07:31:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

